



EC declaration of conformity

For the following equipment :

Product Name: Switching Power Supplies

Model Designation: MPX-120Y (X=S,D,T,Q ; Y=-3.3,-5,-12,-15,-24,-48,A,B,C,D,E)

is herewith confirmed to comply with the requirements set out in the Council Directive 93/42/EEC concerning Medical devices, the following standards were applied :

RoHS Directive (2011/65/EU)

Low Voltage

EN60601-1:1990+A1+A2+A13

TUV certificate No : TA50053140

EMI (Electro-Magnetic Interference)

Conducted emission / Radiated emission

EN55011:2009+A1:2010

Class B

EN55022:2010

Class B

EN61000-6-3:2007+A1:2011

Harmonic current

EN61000-3-2:2006+A1:2009+A2:2009

Voltage flicker

EN61000-3-3: 2008

EMS (Electro-Magnetic Susceptibility)

EN60601-1-2:2007 EN55024:2010 EN61204-3:2000 EN61000-6-2:2005

ESD air

EN61000-4-2:2009

Level 3

8KV

ESD contact

EN61000-4-2:2009

Level 2

6KV

RF field susceptibility

EN61000-4-3: 2006+A1:2008+A2:2010

Level 3

10V/m

EFT bursts

EN61000-4-4:2012

Level3

2KV/5KHz

Surge susceptibility

EN61000-4-5:2006

Level 4

2KV/Line-Line 4KV/Line-Earth

Conducted susceptibility

EN61000-4-6:2009

Level 3

10V

Magnetic field immunity

EN61000-4-8: 2010

Level 4

30A/m

Voltage dip, interruption

EN61000-4-11:2004 >95% dip 0.5 periods 30% dip 25 periods >95% interruptions 250 periods

Note:

A component power supply with load will be installed into final equipment which consists of an electronically shielded metal enclosure. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure.

For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on <http://www.meanwell.com>)".

This Declaration is effective from serial number EB4xxxxxxx

Person responsible for marking this declaration :

Mean Well Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan (R.O.C.)

(Manufacturer Address)

Johnny Huang/Senior Verification Engineer :

(Name / Position)

Johnny Huang
(Signature)

Ted Cheng/Product Manager :

(Name / Position)

Ted Cheng
(Signature)

Taiwan

(Place)

Feb. 27th, 2014

(Date)